



Advanced Test Equipment Rentals
www.atecorp.com 800-404-ATEC (2832)

PROGRAMMABLE NOISE
GENERATING INSTRUMENT
NC 7110
OPERATING MANUAL

Noise Com, Inc.
E64 Midland Ave.
Paramus, NJ 07652

RATION

Note: This is a generic NC7110 manual. Your instrument may differ with some of the options. However, this manual covers all that's needed to operate your instrument regardless of any options. For more questions please call our customer support at (201)261-8797.

II. SPECIFICATIONS

2-1 INSTRUMENT SPECIFICATIONS

Frequency Range 100 Hz - 500MHz

Output Power 10 Mw

Output impedance 50 Ohms.

Input voltage 115 Volts 60Hz (100 & 230V 50Hz Optional).

Noise output is white gaussian noise with a qualified peak factor.

Control by means of front panel or remotely by means of a controller via IEEE (GPIB), RS-232, RS-422 or RS-423.